

SOLIScript® 1-step CoV Kit

Data Sheet

Cat. No.	Size (20 µl Reactions)
08-65-0000S	50 reactions
08-65-00250	250 reactions
08-65-05000	5000 reactions

Description:

SOLIScript® 1-step CoV Kit is optimized for highly sensitive one-step real-time RT-PCR (RT-qPCR) detection of viral RNA of severe acute respiratory syndrome coronavirus 2 (SARS-CoV-2) from RNA samples extracted from swab specimens. Specimen processing should be performed in accordance with pertaining national biological safety regulations and following the recommended World Health Organization (WHO) guidelines on biosafety and biosecurity.

SOLIScript® 1-step CoV Kit is suitable for probe-based detection of up to four targets simultaneously.

NB! SARS-CoV-2 primers and probes, applicable positive and internal extraction controls are not provided with the Kit and should be supplied by the user.

Kit Content:

Component	Size		
	50 rxn	250 rxn	5000 rxn
40x One-step SOLIScript® CoV Mix	25 µl	125 µl	2.5 ml
5x One-step Probe CoV Mix	200 µl	1.0 ml	20.0 ml
Nuclease-free water	1.25 ml	5.0 ml	100.0 ml

Mix compositions:

- 40x One-step SOLIScript® CoV Mix: SOLIScript® Reverse Transcriptase, RiboGrip™ RNase Inhibitor
- 5x One-step Probe CoV Mix: HOT FIREPol® DNA polymerase, reaction buffer, dNTPs, 15 mM MgCl₂ (1x qPCR solution - 3 mM MgCl₂)

Compatible real-time instruments:

The Kit is compatible with ROX-independent and ROX dependent qPCR platforms. NOTE: If using ROX-dependent qPCR cyler, ROX channel should be deactivated and set to "none" while selecting passive reference that will be used in the assay.

Shipping and Storage conditions:

Routine storage: -20 °C. Shipping and temporary storage for up to 1 week at room temperature.

Recommended RT-qPCR reaction mix:

Stock and final concentrations of primers and probe(s) are dependent on detection panel used in the assay.

Component	Volume	Final conc.
5x One-step Probe CoV Mix	4 µl	1x
40x One-step SOLIScript® CoV Mix	0.5 µl	1x
Forward Primer(s)	Variable	200-600 nM
Reverse Primer(s)	Variable	200-800 nM
Probe(s)	Variable	100-250 nM
RNA template	5 µl	
Nuclease-free water	up to 20 µl	
Total	20 µl	

NOTE: For multiple reactions, preparation of a master mix of common components is crucial to reduce pipetting errors. Scale all components according to sample number and reaction volumes, plus add 10% extra volume in order to accommodate pipetting errors.

Recommended RT-qPCR cycling protocol:

Optimal thermocycling protocol depends on a particular set of primers and probe(s) used in the molecular assay.

Step	Temp.	Time	
Reverse transcription	48–55 °C	10–30 min	
Initial activation ¹	95 °C	10 min	
Denaturation	95 °C	5 sec	45 cycles
Annealing/elongation	55–60 °C	30 sec	

¹ 10 min initial activation is crucial for a full activation of HOT FIREPol® DNA Polymerase

NB! The Kit has been tested and validated by diagnostic laboratories using primer-probe panels recommended by Charité, (Germany) and CDC (USA) protocols according to WHO "Coronavirus disease (COVID-19) technical guidance: Laboratory testing for 2019-nCoV in humans", section: "Molecular assays to diagnose COVID-19"; subsection: "In-house developed molecular assays".

Specifically adjusted and technically validated protocols for Charité (Germany) and CDC (USA) diagnostic assays can be found in Supplementary Data Sheet.

For Research Use Only. This product alone does not provide any diagnostic result.

Other recommendations:

To avoid repeated freezing and thawing as well as to minimize the contamination risk of stock solutions of reagents, it is highly recommended to divide large-volume stocks into several smaller aliquots and store them at -20 °C.

Download Supplementary Data Sheet:**More resources available here:**

https://www.solisbiodyne.com/EN/products/onestep_rtqpcr/soliscript_1step_cov_kit

Safety warnings and precautions:

This product and its components should be handled only by persons trained in laboratory techniques. It is advisable to wear suitable protective clothing, such as laboratory overalls, gloves and safety glasses. Care should be taken to avoid contact with skin or eyes. In case of contact with skin or eyes, wash immediately with water. See Safety Data Sheet for additional information.

Revised 06.04.2020

Permitted Use: This product is supplied for Research Use Only (the **Permitted Use**). It is suitable for use as a component of molecular diagnostic assays, where applicable country laws allow. This product alone does not provide any diagnostic result. If the customer wishes to use the product for any purpose other than the Permitted Use, including (without limitation) resale or alteration, the customer should obtain the appropriate license from Solis BioDyne. Some applications of this product may require a license/licenses from one or more third parties which are not provided by the purchase of this product. Users should obtain the license if required. Covered by the patent EP2501716, made by the methods of US Patent No 9,321,999.

Warranty and Disclaimer: This product shall comply with its relevant specification and be fit for its stated purpose, but Solis BioDyne gives no other warranty and makes no representation as to description or quality. Any such warranty or representation is excluded, to the fullest extent permitted by law. In particular, but without limiting the foregoing, Solis BioDyne shall not be liable for the failure of the product to comply with its relevant specification where such failure arises as a result of: (i) customer negligence or because the customer failed to follow any of the applicable technical data or safety sheets, standard user materials, use guidelines or any other information provided by Solis BioDyne as to the storage, transportation, handling, use or maintenance of the products or other good practice regarding the same, or (ii) the customer altering the products in any way without the prior written consent from Solis BioDyne, or (iii) the products differing from the relevant specification as a result of changes made to ensure their compliance with applicable statutory or regulatory requirements.

Nothing shall limit or exclude Solis BioDyne's liability for death or personal injury caused by its negligence, fraud or fraudulent misrepresentation or any matter in respect of which it would be unlawful for Solis BioDyne to exclude or restrict liability. Without limiting the foregoing, Solis BioDyne shall under no circumstances whatever be liable to the customer, whether in contract, tort (including negligence), breach of statutory duty, or otherwise, for any loss of profit, or any indirect or consequential loss arising under or in connection with the products and Solis BioDyne's total liability to the customer in respect of all other losses arising under or in connection with the product, whether in contract, tort (including negligence), breach of statutory duty, or otherwise, shall in no circumstances exceed the price of the products supplied in respect of which the liability has arisen.

HOT FIREPoJ® and SOLIScript® are EU registered trademarks of Solis BioDyne OÜ. RiboGrip™ is a trademark of Solis BioDyne OÜ.

SOLIScript® 1-step CoV Kit

Supplementary Data Sheet

Specifically adjusted protocols for Charité (Germany)^{1,2}
and CDC (USA)³ diagnostic assays

Revised 06.04.2020

General notes:

1. The Kit has been tested and validated by diagnostic laboratories using primer-probe panels recommended by Charité, (Germany)^{1,2} and CDC (USA)³ protocols according to WHO "Coronavirus disease (COVID-19) technical guidance: Laboratory testing for 2019-nCoV in humans", section: "Molecular assays to diagnose COVID-19"; subsection: "In-house developed molecular assays".
2. SARS-CoV-2 primers and probes, applicable positive and internal extraction controls are not provided with the Kit and should be supplied by the user. Being listed in this document does not imply any endorsement or scientific and diagnostic validation by Solis BioDyne.
3. To avoid repeated freezing and thawing as well as to minimize the contamination risk of stock solutions of reagents, it is highly recommended to divide large-volume stocks into several smaller aliquots and store them at -20 °C.
4. For multiple reactions, preparation of a master mix of common components is crucial to reduce pipetting errors. Scale all components proportionally according to sample number and reaction volumes. Make sure you use enough of each reagent for your reactions, plus 10% extra volume to accommodate pipetting errors.

Thermal conditions recommended for Charité (Germany) and CDC (USA) diagnostic assays

Thermal conditions compatible with most cyclers:

Step	Temp.	Time	
Reverse transcription	55 °C	10 min	
Initial denaturation	95 °C	10 min	
Denaturation	95 °C	15 sec	45 cycles
Annealing/Elongation	58 °C	30 sec	

Alternative thermal conditions:

Step	Temp.	Time	
Reverse transcription	48 °C	30 min	
Initial denaturation	95 °C	10 min	
Denaturation	95 °C	5 sec	45 cycles
Annealing/Elongation	55 °C	30 sec	

¹ Corman et al. 2020 „Detection of 2019 novel coronavirus (2019-nCoV) by real-time RT-PCR“, doi: 10.2807/1560-7917.ES.2020.25.3.2000045

² Adapted from https://www.who.int/docs/default-source/coronaviruse/whoinhouseassays.pdf?sfvrsn=de3a76aa_2

³ CDC 2019-Novel Coronavirus (2019-nCoV) Real-Time RT-PCR Diagnostic Panel (<https://www.fda.gov/media/134922/download>)

1. Diagnostic detection of SARS-CoV-2 (2019-nCoV) using Charité, Berlin Germany protocol

Assay/Use	Oligonucleotide ID	Sequence (5'–3')	Comment
RdRP gene	RdRP_SARSr-F2	GTGARATGGTCATGTGTGGCGG	Use 600 nM per reaction
	RdRP_SARSr-R1	CARATGTAAASACACTATTAGCATA	Use 800 nM per reaction
	RdRP_SARSr-P2	FAM-CAGGTGGAACCTCATCAGGAGATGC-BBQ	Specific for 2019-nCoV, will not detect SARS-CoV. Use 100 nM per reaction and mix with P1
	RdRP_SARSr-P1	FAM-CCAGGTGGWACRTCATCMGGTGATGC-BBQ	Pan Sarbeco-Probe, will detect 2019-nCoV, SARS-CoV and bat-SARS-related CoVs. Use 100 nM per reaction and mix with P2
E gene	E_Sarbeco_F1	ACAGGTACGTTAATAGTTAATAGCGT	Use 400 nM per reaction
	E_Sarbeco_R2	ATATTGCAGCAGTACGCACACA	Use 400 nM per reaction
	E_Sarbeco_P1	FAM-ACACTAGCCATCCTTACTGCGCTTCG-BBQ	Use 200 nM per reaction

1.1 Reaction Mix for First line screening assay (E assay)

Component	Volume	
Nuclease-free water	8.5 µl	
5x One-step Probe CoV Mix	4 µl	
40x One-step SOLIScript® CoV Mix	0.5 µl	
Primer E_Sarbeco_F1 (10 µM stock solution)	0.8 µl	ACAGGTACGTTAATAGTTAATAGCGT
Primer E_Sarbeco_R2 (10 µM stock solution)	0.8 µl	ATATTGCAGCAGTACGCACACA
Probe E_Sarbeco_P1 (10 µM stock solution)	0.4 µl	FAM-ACACTAGCCATCCTTACTGCGCTTCG-BBQ
Template RNA, add	5 µl	
Total reaction volume	20 µl	

1.2 Reaction Mix for Confirmatory assay (RdRP assay; detect 2019-nCoV, SARS-CoV and bat-SARS-related CoVs)

Component	Volume	
Nuclease-free water	7.3 µl	
5x One-step Probe CoV Mix	4 µl	
40x One-step SOLIScript® CoV Mix	0.5 µl	
RdRP_SARSr-F2 (10 µM stock solution)	1.2 µl	GTGARATGGTCATGTGTGGCGG
RdRP_SARSr-R1 (10 µM stock solution)	1.6 µl	CARATGTTAAASACACTATTAGCATA
RdRP_SARSr-P1 (10 µM stock solution)	0.2 µl	FAM- CCAGGTGGWACRTCATCMGGTGATGC-BBQ
RdRP_SARSr-P2 (10 µM stock solution)	0.2 µl	FAM-CAGGTGGAACCTCATCAGGAGATGC- BBQ
Template RNA, add	5 µl	
Total reaction volume	20 µl	

1.3 Reaction Mix for Discrimatory assay (RdRp assay, specific for 2019-nCoV)

Component	Volume	
Nuclease-free water	7.5 µl	
5x One-step Probe CoV Mix	4 µl	
40x One-step SOLIScript® CoV Mix	0.5 µl	
RdRP_SARSr-F2 (10 µM stock solution)	1.2 µl	GTGARATGGTCATGTGTGGCGG
RdRP_SARSr-R1 (10 µM stock solution)	1.6 µl	CARATGTTAAASACACTATTAGCATA
RdRP_SARSr-P2 (10 µM stock solution)	0.2 µl	FAM-CAGGTGGAACCTCATCAGGAGATGC- BBQ
Template RNA, add	5 µl	
Total reaction volume	20 µl	

2. Diagnostic detection of SARS-CoV-2 (2019-nCoV) using CDC (US) protocol

Name	Description	Sequence (5'-3')	Working conc.
2019-nCoV_N1-F	2019-nCoV_N1 Forward Primer	5'-GACCCCAAATCAGCGAAAT-3'	20 µM
2019-nCoV_N1-R	2019-nCoV_N1 Reverse Primer	5'-TCTGGTACTGCCAGTTGAATCTG-3'	20 µM
2019-nCoV_N1-P	2019-nCoV_N1 Probe	5'-FAM-ACCCCGCATTACGTTTGGTGGACC-BHQ1-3'	5 µM
2019-nCoV_N2-F	2019-nCoV_N2 Forward Primer	5'-TTACAAACATTGGCCGCAA-3'	20 µM
2019-nCoV_N2-R	2019-nCoV_N2 Reverse Primer	5'-GCGCGACATTCCGAAGAA-3'	20 µM
2019-nCoV_N2-P	2019-nCoV_N2 Probe	5'-FAM-ACAATTTGCCCCAGCGCTTCAG-BHQ1-3'	5 µM
RP-F	RNAse P Forward Primer	5'-AGATTTGGACCTGCGAGCG-3'	20 µM
RP-R	RNAse P Reverse Primer	5'-GAGCGGCTGTCTCCACAAGT-3'	20 µM
RP-P	RNAse P Probe	5'-FAM-TTCTGACCTGAAGGCTCTGCGCG-BHQ1-3'	5 µM

Reaction Mix for 2019-nCoV_N1, 2019-nCoV_N2 and RP singleplex assays

Component	Volume	Final concentration
Nuclease-free water	9 µl	
5x One-step Probe CoV Mix	4 µl	1x
40x One-step SOLIScript® CoV Mix	0.5 µl	1x
Primer F (20 µM stock solution) ¹	0.5 µl	500 nM
Primer R (20 µM stock solution) ¹	0.5 µl	500 nM
Probe (5 µM stock solution) ¹	0.5 µl	125 nM
Template RNA, add	5 µl	
Total reaction volume	20 µl	

¹ If a Combined Primer/Probe Mix is used, pipet 1.5 µl of Combined Primer/Probe Mix per reaction.